

## PROJECT NARRATIVE

Project: Case #416

Pace Project No.: 30207556

---

Date: January 18, 2017

The reported results are on an "as-received" basis. Following the client's specifications, samples were analyzed directly in the containers received. The dimensions of the sampling containers were different than those Pace uses for gamma spectroscopy calibration purposes. The results reported should be construed as estimated due to geometric differences between the sample sources and Pace's 8-ounce can gamma geometry.

## REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Pace Analytical Services, LLC.